



CERTIFICATE OF CALIBRATION

Calibration Certificate #2046.01

10041 **7632B Hub Parkway Valley View, OHIO 44125** **216-525-0050**

NAME:	Rigol Technologies	PURCHASE OR.	RU090090
ADDRESS:	7401 First Place Suite N	LABOR	
CITY/STATE:	Oakwood Village OH	MATERIAL	
ZIP:	44146	TOTAL COST	
CONTACT:	Hossam	ON-SITE	NO
REC'D COND.	Within Spec	RETURNED	Within Spec
MANUFACTURER:	Rigol	CAL. DATE	7-Apr-11
MODEL:	DG1022 ary Waveform Gen	CAL DUE:	7-Apr-12
SERIAL / ASSET:	DG1D130400480 NA	TEMP/ RH%	23.8°C / 37.3%

	AS FOUND	DEVIATION	AS LEFT	DEVIATION
Frequency Synthesis Accuracy: ± 20 ppm				
MHz	20.00	19.99986	-7 ppm	19.99986
ch2	20.00	19.99988	-6 ppm	19.99988
Amplitude Accuracy, 50 Ohms: ± 1% Setting + .707 mV				
mVrms	10.00	9.92200	-0.780%	9.92200
	100.00	99.8461	-0.154%	99.8461
	500.00	500.103	0.021%	500.103
Vrms	1.00	0.994960	-0.504%	0.994960
	3.53	3.51922	-0.305%	3.51922
Amplitude Accuracy, High-Z: ± 1% Setting + .707 mV				
mVrms	15.00	14.7963	-1.358%	14.7963
	100.00	99.9640	-0.036%	99.9640
	500.00	497.937	-0.413%	497.937
Vrms	1.00	0.998919	-0.108%	0.998919
	5.00	4.97655	-0.469%	4.97655
DC Offset: ± 2% Setting				
mVDC	10.00	9.9050	-0.95%	9.9050
	100.00	98.434	-1.57%	98.434
	500.00	501.502	0.30%	501.502
VDC	1.00	1.01066	1.07%	1.01066
	5.00	5.01465	0.29%	5.01465

EQUIPMENT USED IN CALIBRATION:				BIS TEST REPORT NO: 406405914
QC#	Manufacturer	Model	Date Due	Traceability information: This calibration was conducted using an unbroken chain of standards to National Institute of Standards and Technology.
0488	AGILENT	34401A	4-May-11	
0203	RACAL-DANA	01992	15-May-11	
0648	OMEGA	I SERVER	12-May-11	
0699	TEKTRONIX	TPS 2012	5-Oct-11	
0648	OMEGA	I SERVER	12-May-11	

Customer approval: Reviewed By: Jeffery P. Metrologist **Brian Tussel**

All equipment used for this calibration is traceable to National Standards (e.g. NIST). Calibration meets ISO/IEC 17025, The Uncertainty of standards utilized for calibration meet a Test Uncertainty Ratio (TUR) of 4:1 (at k=2, 95% confidence interval) or better unless otherwise stated.

While the calibration due date is recommended, the client is advised that any number of factors may cause the item to drift out of calibration before the recommended due date expires. It is the responsibility of the client to monitor the calibration status while using the equipment



CERTIFICATE OF CALIBRATION

Calibration Certificate #2046.01

7632B Hub Parkway Valley View, OHIO 44125

216-525-0050

MANUFACTURER	<i>Rigol</i>			
MODEL	<i>DG1022</i>			
SERIAL NUMBER	<i>DG1D130400480</i>	ASSET	<i>NA</i>	
	AS FOUND	% DEVIATION	AS LEFT	% DEVIATION
Amplitude Flatness:	<i>±1% < 10 MHz, ±1.5% 10-25 MHz, ±4% >25 MHz</i>			
1 kHz (Ref.)	0.504	<i>Ref.</i>	0.504	<i>Ref.</i>
100 Hz	0.504	0.00%	0.504	0.00%
10 kHz	0.508	0.79%	0.508	0.79%
100 kHz	0.508	0.79%	0.508	0.79%
1 MHz	0.504	0.00%	0.504	0.00%
10 MHz	0.504	0.00%	0.504	0.00%
20 MHz	0.500	-0.79%	0.500	-0.79%
Modulation:	Within Spec		Within Spec	
Sweep Functions:	Within Spec		Within Spec	
Waveshapes:	Within Spec		Within Spec	
BIS TEST REPORT NO:			406405914	
While the calibration due date is recommended, the client is advised that any number of factors may cause the item to drift out of calibration before the recommended due date expires. It is the responsibility of the client to monitor the calibration status while using the equipment				
FORM #10C Dec '07 REV.1.10 Note this form may not be reproduced except in full without written permission of Broadview Instrumentation Service. Inc.				